

MEDICAL POLICY

Medical Policy Title	External Prosthetic Devices
Policy Number	1.01.18
Current Effective Date	August 17, 2026
Next Review Date	April 2027

Our medical policies are guides to evaluate technologies or services for medical necessity. Criteria are established through the assessment of evidence based, peer-reviewed scientific literature, and national professional guidelines. Federal and state law(s), regulatory mandates and the member's subscriber contract language are considered first in the determination of a covered service. (Link to [Product Disclaimer](#))

This policy addresses external prosthetics only. Please refer to specific policies for implantable prosthetic devices and orthotic devices.

POLICY STATEMENT(S)

General Criteria

- I. External prosthetic devices are considered **medically appropriate** to replace all or part of an internal organ or replace the function of a permanently inoperative or malfunctioning body part. (Refer to [Lower Extremity](#) and [Upper Extremity](#) sections regarding specific criteria.)
- II. Supplies needed to make a covered, medically appropriate external prosthetic device functional are considered **medically appropriate**. Examples of covered supplies include tracheostomy kits, ostomy supplies, urine pouches, and batteries to operate an artificial larynx.
- III. Back-up prosthetic devices are considered **not medically necessary**; more than one prosthetic device is considered a matter of convenience for the member.
- IV. Prosthetics used for activities other than normal activities of daily living, including, but not limited to, those utilized for leisure or sporting activities such as skiing or swimming are considered **not medically necessary**.
- V. Devices or implants used primarily for cosmetic purposes are considered **not medically necessary**.
- VI. Osseointegration surgery and/or prosthesis for upper and/or lower limbs (e.g., Osseoanchored Prostheses for the Rehabilitation of Amputees (OPRA) Implant System, Axor II connection device) are considered **investigational**.

Preparatory Prosthesis

- VII. Preparatory prostheses are considered **medically appropriate** after surgery to prevent edema of the residual limb and to enhance functional recovery.
- VIII. Additions such as protective covers, ultralight material, nonstandard components (e.g., microprocessor knees), and flex foot systems (e.g., energy storing) are considered not medically necessary for a preparatory prosthesis.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 2 of 24

Lower Extremity Prosthesis

- IX. Basic preparatory or permanent (definitive) lower limb prosthetic devices are considered **medically appropriate** for individuals with a function level 1 or greater including but not limited to the following: (Refer to the Description section for definitions of [functional levels](#).)
- A. A single axis, constant friction knee and other basic knee systems for persons whose functional level is 1 or above;
 - B. A fluid or pneumatic knee for persons whose functional level is 3 or above;
 - C. A high-activity knee control frame for members whose function level is 4;
 - D. Solid-ankle cushioned-heel (SACH) foot (L5970) or single axis ankle/foot (L5974) appropriate for sedentary patients with a functional level 1 or above;
 - E. Flexible-keel foot (L5972) or multi-axial ankle/foot (L5978) for a functional level of 2 or above;
 - F. Energy storing foot (L5976), dynamic response foot with multi-axial ankle (L5979), flex foot system (L5980), flexwalk system or equal (L5981) or shank foot system with vertical loading pylon (L5987) for a functional level of 3 or above;
 - G. A prosthetic shoe as a terminal device to supplement a substantially absent foot (L3250).
- X. A socket for a lower limb prosthetic device is considered **medically appropriate** for **ANY** of the following:
- A. One (1) socket per individual prosthetic;
 - B. Two (2) of the same socket inserts per individual prosthesis at the same time;
 - C. Up to two (2) test sockets when fitting for a socket;
- A test socket is not recommended for immediate postsurgical or early fitting prostheses.
- XI. A vacuum-assisted socket system (VASS) for a lower limb prosthetic device is considered **medically appropriate** for **ANY** of the following situations:
- A. There is a nonhealing skin breakdown on the stump from friction due to an ill-fitting socket;
 - B. The current socket can no longer be modified to adequately secure the limb to the prosthesis.
- XII. A microprocessor-controlled knee prosthesis (e.g., L5856) is considered **medically appropriate** when **ALL** of the following criteria are met:
- A. For persons with a knee disarticulation amputee **OR** transfemoral amputee;
 - B. Functional level 3 or above;
 - C. Lower-level prosthetic devices (e.g., fluid, pneumatic, etc.) cannot be used or are insufficient to meet the functional needs of the individual in performing ADLs;
 - D. The patient has adequate cardiovascular and pulmonary reserve;
 - E. The patient has received additional training for use of this technology and has demonstrated

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 3 of 24

adequate cognitive ability to master use and care requirements;

- F. The patient demonstrated a need to perform **ALL** of the following:
 - 1. long distance ambulation at variable rates;
 - 2. regular and frequent ambulation on uneven terrain (e.g., grass, gravel, or curbs); **and**
 - 3. regular and frequent ambulation on stairs or ramps;
- G. Functional evaluation indicates that with training, use of a microprocessor-controlled knee prosthesis is likely to meet the functional needs of the individual.
- H. Microprocessor-controlled knees are **contraindicated** for **ANY** of the following:
 - 1. The patient's [functional level](#) is less than 3 or has limited ambulation due to poor balance or ataxia;
 - 2. The patient is unable to tolerate the weight of the prosthesis;
 - 3. The patient is unable to use the swing and stance features of the knee;
 - 4. The patient is unable to change the prosthesis or has a condition that would cause inadequate fitting;
 - 5. The patient has significant hip flexion contracture (over 20 degrees);
 - 6. The patient has significant deformity of the remaining limb that would impair ability to stride; **or**
 - 7. The prosthesis will be used when the environmental conditions include excessive moisture or dust which invalidates the warranty.

XIII. A microprocessor-controlled ankle-foot prosthesis (L5973) (e.g., Proprio Foot, Ossur, Aliso Viejo, CA) is considered **medically appropriate** when **ALL** of the following criteria are met:

- A. For persons with a transtibial amputation;
- B. [Functional level](#) of 3 or above;
- C. Lower-level prosthetic devices (e.g., SACH foot) cannot be used or are insufficient to meet the functional needs of the individual in performing ADLs;
- D. The patient has adequate cardiovascular and pulmonary reserve;
- E. The patient has received additional training for use of this technology and has demonstrated adequate cognitive ability to master use and care requirements;
- F. The patient demonstrated a need to perform **ALL** of the following:
 - 1. long distance ambulation at variable rates;
 - 2. regular and frequent ambulation on uneven terrain (e.g., grass, gravel, or curbs); **and**
 - 3. regular and frequent ambulation on stairs or ramps;

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 4 of 24

- G. Functional evaluation indicates that with training, use of a microprocessor-controlled ankle-foot prosthesis is likely to meet the functional needs of the individual.

XIV. The following lower extremity prosthetics or components are considered **not medically necessary**:

- A. Lower limb prosthetic devices for individuals with [functional level 0](#);
- B. A microprocessor-controlled ankle-foot with power assist (L5973 AND L5969) (e.g., PowerFoot BiOM).

Upper Extremity Prosthesis

XV. A conventional body-powered, upper extremity prosthesis is considered **medically appropriate** to replace all or part of an upper extremity or replace the function of a permanently inoperative or malfunctioning upper extremity.

XVI. Myoelectric upper arm prosthetic components are considered **medically appropriate** when **ALL** of the following are met:

- A. Standard body-powered prosthetic devices cannot be used or are insufficient to meet the functional needs of the individual in performing ADLs;
- B. The remaining musculature of the arm(s) contains the minimum microvolt threshold to allow operation of a myoelectric prosthetic device;
- C. The patient is free of comorbidities that could interfere with function of the prosthesis (neuromuscular disease, etc.);
- D. The patient has demonstrated sufficient physiological and cognitive function to allow effective operation of a myoelectric prosthetic device;
- E. Functional evaluation indicates that with training, use of a myoelectric prosthesis is likely to meet the functional needs of the individual (e.g., gripping, releasing, holding, and coordinating movement of the prosthesis) when performing ADLs. This evaluation should consider the patient's needs for control, durability (maintenance), function (speed, work capability), and usability.

XVII. The following upper extremity prosthetics are considered **investigational**:

- A. Myoelectric hand (including partial hand) with individual control of digits (e.g., Michelangelo hand [OttoBock], AxonArm Ergo [OttoBock], ProDigits);
- B. Enhanced dexterity upper-limb prosthetic components with both sensor and myoelectric control (e.g., LUKE Arm).

Device Repair

XVIII. Repair of a medically necessary external prosthetic or components not under warranty will be considered **medically appropriate** when the following criteria are met:

- A. Physician documentation includes **ALL** of the following:
 - 1. date of device implantation/initiation;

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 5 of 24

2. manufacturer warranty information, if applicable;
 3. attestation that the patient has been compliant with the use of device and will continue to benefit from the use of device;
- B. The device is no longer functioning adequately; and **BOTH** of the following criteria are met:
1. inadequate function interferes with activities of daily living; **and**
 2. repair is expected to make the equipment fully functional (as defined by manufacturer).
- XIX. Repair of equipment damaged due to patient neglect, theft, abuse, or when another available coverage source is an option (e.g., homeowners, rental, auto, liability insurance, etc.) is **ineligible for coverage**.

Device Replacement

- XX. Replacement of a medically necessary external prosthetic or components not under warranty will be considered **medically appropriate** when **EITHER** of the following criteria are met:
- A. The device is no longer functioning adequately and has been determined to be non-repairable, or the cost of the repair is in excess of the replacement cost;
 - B. There is documentation that a change in the patient's condition makes the present unit non-functional and improvement is expected with a replacement unit.
- XXI. The replacement of a properly functioning external prosthetic, its components or accessories is considered **not medically necessary**. This includes, but is not limited to, replacement desired due to advanced technology or in order to make the device more aesthetically pleasing.
- XXII. The replacement of equipment damaged or lost due to patient neglect, theft, abuse, or when another available coverage source is an option (e.g., homeowners, rental, auto, liability insurance, etc.) is **ineligible for coverage**.
- XXIII. Accessories or components for external prosthetics that are considered not medically necessary or investigational by peer-reviewed literature will also be considered as **not medically necessary** or **investigational** by the Health Plan.

RELATED POLICIES

Corporate Medical Policy

1.01.00 Durable Medical Equipment –Standard and Non-Standard

1.01.25 Orthotics

7.01.30 Erectile Dysfunction

10.01.01 Breast Reconstruction Surgery and Prophylactic Breast Cancer Risk-Reducing Mastectomy

POLICY GUIDELINE(S)

- I. Documentation Requirements include **ALL** of the following:

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 6 of 24

- A. General description of the prosthetic including but not limited to a HCPCS code, a HCPCS code narrative, and a brand name, make, model, and description of the requested device features;
 - B. Description of all additional ordered options, accessories, or features that are separately billed or require an up-graded code (List each separately and include quantity);
 - C. Specify whether the prosthetic is an initial, replacement, preparatory, or definitive, or a request to upgrade.
- II. To be eligible for coverage a prosthetic device must address a problem in which the device is needed for at least 90 days.
- III. Coverage of ostomy equipment and supplies is required under the New York Insurance Law, subject to applicable cost-sharing (copayments, deductibles, and/or coinsurance), when the equipment and/or supplies are prescribed by a physician or any other health care provider legally authorized to prescribe under Title VIII of the New York Education Law.
- IV. Polishing and resurfacing of an eye prosthesis (V2624) may be performed up to two (2) times per year.
- V. Synthetic wigs are **eligible for coverage** when there is a severe hair loss due to injury, disease, or as a side effect of the treatment of a disease (e.g., chemo/radiotherapy, burns, chronic inflammatory conditions such as alopecia). Wigs made from human hair are not covered unless there is an allergy to all synthetic wig materials.

DESCRIPTION

External prosthetic devices, which are worn as anatomic supplements, are used to replace non-functioning or absent body parts. Examples of external prosthetic devices include artificial limbs, removable artificial eyes, external breast prostheses or prosthetic bras for post mastectomy patients, external pacemakers, and electronic speech aids for post-laryngectomy patients. Some HCPCS "A" code items such as ostomy bags for a patient with an artificial stoma, become prosthetic devices.

The design of lower limb prosthetic devices is based on the functional level classification of the individual as described by Medicare Guidelines.

- I. Level 0: Does not have the ability or potential to ambulate or transfer safely with or without assistance and a prosthesis does not enhance quality of life or mobility.
- II. Level 1: Has the ability or potential to use a prosthesis for transfers or ambulation on level surfaces at fixed cadence. Typical of the limited and unlimited household ambulator.
- III. Level 2: Has the ability or potential for ambulation with the ability to traverse low level environmental barriers, such as curbs, stairs, or uneven surfaces. Typical of the limited community ambulator.
- IV. Level 3: Has the ability or potential for ambulation with variable cadence. Typical of the community ambulator who has the ability to traverse most environmental barriers and may have vocational, therapeutic, or exercise activity that demands prosthetic utilization beyond simple locomotion.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 7 of 24

- V. Level 4: Has the ability or potential for prosthetic ambulation that exceeds basic ambulation skills, exhibiting high impact, stress, or energy levels. Typical of the prosthetic demands of the child, active adult, or athlete.

Lower limb prosthetic devices are either preparatory or permanent (definitive). A preparatory prosthesis is a temporary device fitted while the residual limb is still remolding after surgery. The preparatory prosthesis is used until the residual limb has reached its final shape and size, typically within three to six months. Once the residual limb is stabilized (e.g., the residual limb volume is unchanged, and the socket fit is consistent for two to three weeks) a permanent or definitive prosthesis can be fitted.

A basic preparatory or permanent (definitive) lower limb prosthetic device consists of the following components: 1) socket, 2) suspension mechanism, 3) knee joint, 4) pylon, and 5) terminal device (foot). Described below are definitions of each component and usual indications. The listing is not all-inclusive.

1. Socket: the interface between the residual limb and the prosthesis, functions to protect the residual limb and transmits the forces associated with ambulation and standing. They can be soft made of foam, rubber or leather, or hard made of acrylic or thermoplastic. A socket is necessary to secure the safety of the residual limb and provides a rigid control of the prosthesis. The socket should cause minimal discomfort during its usage. Additions, such as liners, sleeves, and socks to provide improved fit of the socket to the residual limb.
2. Suspension mechanism: method which holds the prosthesis to the body. There are several types which include locking pin, TES belt, suspension sleeve, waist belt, suction, and vacuum. A shuttle lock/pin comprised of a liner with a pin placed into the end and a locking mechanism. The liner improves contact between the limb and the prosthesis. The pin improves suspension from the deficient limb. A Silesian belt fastens to the socket laterally, above the greater trochanter, and wraps around the opposite iliac crest. A Silesian belt is appropriate for the pediatric patient as well. The gel liner suction system uses a gel elastomeric liner, and a pin may or may not be used. Gel liner suction system is appropriate for patients with a transfemoral or transtibial amputation. Standard suction contains a one-way air valve in the distal end; air is expelled after the socket is donned creating a seal from the development of a small negative pressure. Standard suction is a common suspension choice for transfemoral prostheses. Vacuum suspension is created between an airtight sleeve and a one-way air valve located in the bottom of the socket. Vacuum suspension is another transtibial suspension option. Lastly, the vacuum-assisted socket system (VASS) works by use of a vertical shock pylon that acts as a vacuum pump and continually withdraws air from the sealed socket during ambulation. There is insufficient evidence to support the efficacy of vacuum-assisted socket systems (VASS) over standard socket types for all patients, however VASS may be appropriate for individuals where the current socket can no longer be modified to adequately secure the limb to the prosthesis or there is non-healing skin breakdown on the stump due to an ill-fitting socket.
3. Knee joint: provides support during the stance phase of ambulation, produces smooth control during the swing phase, and maintains unrestricted motion for sitting and kneeling. There are several types of knee joints including single-axis knees, polycentric-axis knees, hydraulic

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 8 of 24

knees, and microprocessor-controlled knees. Single-axis knees - recommended for: classification level 1 or above. Polycentric-axis knees -recommended for: classification level 3 or above. Hydraulic knees are chosen by more active amputees. Hydraulic knees - recommended for: classification level 3 or above. A microprocessor-controlled knee (e.g., Otto Bock C-leg, Otto Bock Genium X3) are single- or multi-axial energy saving knee with onboard microprocessor. This allows the knee to adjust for variable gait cycles providing more natural movement during stair descent or while ambulating on uneven terrain.

4. Pylon: attaches the socket to the terminal device. Allows axial rotation and is able to absorb, store, and release energy.
5. Terminal device (foot): functions to provide a stable, weight-bearing surface, absorb shock, replace lost muscle function, replicate the anatomic joint, and to restore cosmetic appearance. There are several types of terminal devices including non-energy, energy- returning and microprocessor-controlled ankle foot systems. Non-energy devices are the solid-ankle cushioned-heel (SACH) foot and the single-axis foot. The SACH foot is low cost and low maintenance. The single axis foot provides increased knee stability. Either SACH or single axis foot is used in sedentary patients. Energy-returning (energy storing) devices assist the body's natural biomechanics and allow for greater cadence or less oxygen consumption with a multi-axis or dynamic-response. A flexible-keel foot or multi-axial ankle/foot is appropriate for a functional level of 2 or above. An energy storing foot, dynamic response foot with multi-axial ankle, flex foot system, flexwalk system or equal or shank foot system with vertical loading pylon is appropriate for a functional level of 3 or above (Agrawal 2015). The dynamic response foot is the top of the line foot and is commonly used by young, active persons and by athletic individuals. These are made from ultralight materials. Microprocessor-controlled ankle-foot systems use a sensor device (Terrain Logic), which enables the ankle prosthesis to respond appropriately and immediate to variations in ground surface and activity. Examples: Proprio-Foot (Ossur), Meridium (Ottobock), and Elan (Blatchford. Microprocessor-controlled ankle-foot system is considered medically necessary when ADLs cannot be met with standard prosthetic devices. Powered prosthetic ankles, such as the BiOM (Ottobock Empower, Duderstadt, Germany) provide battery-powered mechanical push-off, with the goal of reducing gait asymmetry as well as the energy cost of ambulation for people with transtibial amputations. A prosthetic shoe can function as a terminal device to supplement a substantially absent foot. The function of prosthetic shoes is quite distinct from that of non-covered orthopedic shoes and supportive foot devices, which are used by individuals whose feet, although impaired, are essentially intact.

Upper limb functional prostheses generally can be divided into two categories: body-powered prostheses or externally electrically powered prostheses. Body-powered prostheses are controlled by cables and require gross limb movement. Externally electrically powered prostheses use the electrical activity from select residual limb muscle contractions as a signal to activate the electric motor of the prosthesis using either a myoelectrically controlled or a switch-controlled prostheses. A hybrid system, which is a combination of body-powered and myoelectric components, may be used for high level amputations (at or above the elbow). Hybrid systems allow control of two joints at once (i.e., one body-powered and one myoelectric) and are generally lighter and less expensive than a

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 9 of 24

prosthesis composed entirely of myoelectric components.

Conventional body-powered, upper extremity prostheses have the following components: 1) socket, 2) suspension, 3) control-cable system, 4) terminal device, and 5) components for any interposing joints as needed according to the level of amputation. Described below are definitions of each component. The listing is not all-inclusive.

1. **Socket:** fabricated from lightweight plastic or graphite composite materials. This is composed of a rigid inner socket fit to the residual limb which determines comfort and function, and the outer wall which is the same length and contour as the opposite sound limb.
2. **Suspension system:** holds prosthesis securely to the residual limb; accommodates and distributes the forces associated with the weight of the prosthesis and any superimposed lifting loads. There are several types of systems available including harness-based systems, self-suspending sockets, and suction sockets. Harness based systems are the most commonly used. Self-suspension sockets are limited to wrist or elbow disarticulations and to transradial amputations. They are commonly utilized with an externally powered, myoelectrically controlled, transradial prosthesis. Suction sockets are similar to lower extremity options which contain a one-way air valve in the distal end; air is expelled after the socket is donned creating a seal from the development of a small negative pressure. Suction sockets are appropriate for the patient with a transhumeral amputation.
3. **Control-cable system:** There are 3 types including passive, body-powered, and myoelectric. Passive systems are lightweight, cannot restore function and must be repositioned manually, typically by moving it with the opposite arm. Body-powered systems utilize a harness and cable to provide functional manipulation of the elbow and hand. Voluntary movement of the shoulder and/or residual limb extends the cable and transmits the force to the terminal device. Myoelectric systems use muscle activity from the remaining limb for the control of joint movement. Electromyographic signals from the limb stump are detected by surface electrodes, amplified, and then processed by a controller to drive battery-powered motors that move the hand, wrist, or elbow. These may be considered the most physiologically natural but may be slow and limited to one joint at a time.
4. **Terminal devices:** can be passive or active, voluntary opening (closed at rest), voluntary closing (open at rest), myoelectric hand with individual control of digits, and partial hand prosthesis. There are many different options available for terminal devices depending on occupation, hobbies, or sports. A passive terminal device is more cosmetic than functional and more costly than active terminal devices. An example of a passive terminal device is a child mitt to assist child with crawling. An active terminal device is more functional than cosmetic and can be either a hook or hand. A hand can be powered by cable or external power and is more cosmetically pleasing than a hook. A hook provides active lateral pinch grip. A myoelectric hand with individual control of digits includes the SensorHand by Advanced Arm Dynamics, ProDigits, i-digits quantum, i-limb quantum, i-limb ultra, Michelangelo hand, and AxonArm Ergo, and Select Myoelectric Hand. There is a lack of peer-reviewed literature to evaluate functional outcomes of these myoelectric devices.
5. **Components for any interposing joints** as needed according to the level of amputation.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 10 of 24

A type of advanced upper limb prosthetic device has been developed that incorporates both sensor components and myoelectric control (e.g., Life Under Kinetic Evolution [LUKE] Arm [previously known as the DEKA Arm System, Mobius Bionics, LLC. Manchester, NH] and SensorHand Speed [OttoBock, Austin, TX]) The LUKE Arm was developed in a joint effort between DEKA Research & Development and the U.S. Department of Defense Advanced Research Projects Agency program. It is the first commercially available myoelectric upper limb that can perform complex tasks with multiple simultaneous powered movements. These include up to 10 active degrees of freedom (DOF) or movements, including shoulder abduction, shoulder flexion/extension, humeral rotation, elbow flexion/extension, wrist pronation/supination, wrist flexion/extension, index finger flexion/extension, flexion/extension of other fingers, thumb flexion/extension, and thumb abduction/adduction. In addition to the electromyographic electrodes, the LUKE Arm contains a combination of mechanisms, including switches, movement sensors, and force sensors including grip force feedback. The primary control resides with motion sensors on top of the feet, attached to the user's shoe.

An osseointegrated prosthesis for the rehabilitation of amputees (OPRA), is an implant system for direct skeletal anchorage of amputation prostheses. OPRA constitutes a rehabilitation alternative for transfemoral amputees when treatment with socket prostheses is insufficient. This bone anchored device is intended for skeletally mature individuals who have transfemoral amputation due to trauma or cancer. The OPRA Implant System consists of seven components that are implanted during two surgeries. The overall time commitment for surgery, recovery and rehabilitation is anticipated to be greater than one year. Compared to socket suspension techniques, direct skeletal attachment of a prosthetic limb through osseointegration offers many potential advantages including improved mechanical transfer of motion, reduced skin irritation from a prosthetic socket, improved joint range-of-motion, and enhanced comfort. Osseointegration also presents the risk of serious complications such as infection, failure of the implant, and bone fracture.

SUPPORTIVE LITERATURE

Microprocessor Prosthetic Knees

Thibaut et al (2022) conducted a systematic review including studies of microprocessor prosthetic knees in patients with lower limb amputation. The review included 18 studies (seven (7) RCTs, six (6) cross-sectional studies, and five (5) follow-up studies). Overall, the authors found better functional status and mobility with microprocessor prosthetic knees, but it remains unclear whether there are differences among various models of microprocessor prosthetic knees.

Hahn et al (2022) performed a systematic review and meta-analysis which included 13 published studies of microprocessor prosthetic knees in limited community ambulators. Microprocessor prosthetic knees had improved outcomes in terms of falls, fear of falling, risk of falling, and mobility grade when compared with non-microprocessor prosthetic knees in limited community ambulators.

Alzeer et al (2022) conducted a cross-sectional study with a total of 76 adult unilateral transfemoral amputees classified into two groups. The participants in the first group (38) used the microprocessor-controlled prosthetic knee (Genium, Otto Bock, Minneapolis, MN, USA), and the participants in the second group (38) used various non-microprocessor-controlled prosthetic knee (hydraulic and total knee joints). The microprocessor-controlled prosthetic knee participants showed significantly improved utility, appearance, ambulation, and total Prosthetic Evaluation Questionnaire outcome scores. This

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 11 of 24

study is limited by its small size and observational nature.

Microprocessor-Controlled Ankle-Foot Prostheses

Thomas-Pohl et al (2021) compared three (3) types of microprocessor prosthetic ankles, including the Proprio Foot Ossur, Elan Endolite, and Meridium Ottobock, in a crossover study. The six (6) participants were transtibial amputees usually fit with an energy storing and returning foot. The participants tested all devices for a two (2) week period. The primary outcome was to evaluate the ability of these prostheses to adapt to ground inclination. Overall, the study found that microprocessor prostheses allowed for better posture and a reduction of residual knee moment on positive and/or negative slope when compared to the patients' energy storing and returning feet. This study is limited by its small sample size.

Colas-Ribas et al (2022) conducted a multicenter, randomized, controlled cross-over study in 45 patients with ankle prosthesis in France. Each ankle-foot prosthesis (microprocessor-controlled Proprio Foot or non-microprocessor standard prescribed ankle prosthesis) was worn for a total of 34 days. Energy expenditure was assessed by oxygen uptake measured at the maximum level reached with the two prostheses during treadmill walking at progressively increasing incline and speed. Quality of life and satisfaction was assessed by a questionnaire after wearing each of the two prostheses. The authors found energy expenditure was similar between prostheses (19.4 mL/kg/min with Proprio Foot and 19.1 mL/kg/min with other prostheses) with no statistical difference observed. Quality of life questionnaire physical scores with Proprio Foot were significantly better than with other prostheses (68.5 vs. 62.1; $p=.005$) as were mental scores (72.0 vs. 66.2; $p=.006$).

Weber et al (2022) conducted a retrospective independent analysis to compare differences between non-microprocessor and microprocessor feet (MPF) in the community setting. They measured physical function and mobility outcomes of 20 individuals with transtibial prosthetics that were transitioning to MPF and maintained their original socket fit from July 2019 to February 2021. The most frequently used MPF was Blatchford Elan MPF ($n=12$), Proteor Kinnex, Otto Bock Meridium and Ossur Proprio also utilized. Patient reported outcomes with a Patient Reported Outcomes Measurement Information System- Physical Function (PROMIS-PF) form and a Prosthesis Evaluation Questionnaire (PEQ). Results showed a significant increase in PROMIS-PF scores with MPF feet compared to non MPF feet. All queried items from PEQ (back pain, residual limb pain, and hill ascent and descent) improved with transition to an MPF. Significant improvements noted with prosthesis fit and sitting comfort with the prosthesis, back pain, ascent and descent of steep hills. Mean improvements for residual limb pain intensity did not reach a level of statistical significance, along with perceived weight of the prosthesis. This analysis is limited by a small sample size and difference of MPF models used. The authors concluded this retrospective analysis supports the beneficial impact of MF on improving socket comfort, reducing back pain, improving sit to stand transfers, enhancing hill ascent and descent, and stair negotiation. Additionally, this study demonstrates how MPFs interact outside of the laboratory.

Powered Microprocessor-Controlled Ankle-Foot Prosthesis

Several small crossover comparative studies have been published studying powered microprocessor-controlled ankle-foot prosthetics compared to unpowered prosthetics. Gardinier et al (2018) published a study of ten males with unilateral transtibial amputation who trialed the powered BiOM

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 12 of 24

T2 Ankle System, measuring metabolic costs and preferred walking speed compared to an unpowered prosthesis. There were no significant differences in oxygen consumption (2.9% difference; $P = 0.606$, $d = 0.26$), cost of transport ($\sim 1\%$ difference; $P = 0.652$, $d = 0.23$), or preferred walking speed ($\sim 1\%$ difference; $P = 0.147$, $d = 0.76$) when using the powered ankle compared to unpowered prostheses. Secondary analysis showed that participants with a K4 Medicare Function level were more likely to exhibit energy cost savings than those with lower function levels. The authors concluded the participants did not demonstrate significant improvements in energetics or preferred speed when wearing a clinically tuned powered ankle prosthesis compared to their non-powered prostheses.

Kim et al (2021) conducted a randomized crossover trial comparing the BiOM powered prosthesis to an unpowered device. Ten participants were all male, nine had a K3 Medicare Function level and one participant was a K4 Medicare Function level. Participants' metabolic costs and self-selected walking speeds were measured in the laboratory and daily step count, daily steps away from home, and walking speed were measured over two weeks of at-home prosthesis use. Participants completed the Prosthesis Evaluation Questionnaire (PEQ) and Short Form (SF)-36 after each prosthetic condition and a Prosthesis Preference questionnaire at the end of the study. They found there were no differences in metabolic costs ($p=0.585$), daily step count ($p=0.995$), walking speed in-lab ($p=0.145$) and in daily life ($p=0.226$), or perception of mobility between prostheses ($p\geq 0.058$). With the powered prosthesis, participants had increased self-reported ambulation ($g=0.682$) and decreased frustration ($g=0.506$). The authors concluded there were no universal benefits of the powered prosthesis on function in the lab or home environment. Additionally, self-reported preferences did not often correlate with objective measures of function.

Myoelectric Hand with Individual Digit Control and Partial Hand Prosthesis

Although the availability of a myoelectric hand with individual control of digits has been widely reported in lay technology reports, video clips, and basic science reports, no peer-reviewed publications have been found that evaluate functional outcomes of individual digit control in amputees.

Kerver et al (2023) compared the multi-grip myoelectric hand prosthesis (MHP) to that of a standard myoelectric hand prostheses (SHP) in all categories of the International Classification of Functioning, Disability, and Health-model (ICF-model). MHP users ($N=14$, 64.3% male, mean age=48.6 years) performed physical measurements with their MHP (i-Limb Quantum/Ultra (Touch Bionics; Livingston, United Kingdom), BeBionic (Ottobock; Duderstadt, Germany) or VINCENT (Vincent Systems, Karlsruhe, Germany)) and an SHP to compare the joint angle coordination, dexterity, and prosthetic hand function. In addition, comparisons on user experience, satisfaction and quality of life were performed. The SHP group ($N=19$, 68.4% male, mean age=58.1 years) utilized a Myohand Variplus Speed (Ottobock; Duderstadt, Germany) or Motion Control Hand (Fillauer, USA), which has a movable thumb, index finger, and middle finger that can open and close in only one grip. The authors found no superior benefit of the MHP devices compared to SHP. The SHP scored better than the MHP in several outcome measures. MHP users experienced more pain or limitations due to pain. Limitations of the study include small sample size and lack of randomization.

In a single case study, Widehammar et al (2022) evaluated the effect of a multi-grip myoelectric hand prosthetic performance of daily activities, pain-related disability, and prosthesis use, in

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 13 of 24

comparison with a single-grip myoelectric hand prosthetic. Nine (9) adults participated in the study, and all had previous experience of single-grip myoelectric prostheses and were prescribed a prosthesis with multi-grip functions (Bebionic multi-grip hand, Ottobock, Vienna, Austria). To assess the changes in daily activities, pain-related disability and prosthesis use between single-grip and multi-grip myoelectric prosthetic hands, the Canadian Occupational Performance Measure, Pain Disability Index, and prosthesis wearing time were measured at multiple time-points. At 6 months' follow-up self-perceived performance and satisfaction scores had increased, prosthesis wearing time had increased, and pain-related disability had reduced in participants with musculoskeletal pain at baseline. There was some missing data of the performance tests due to the breakdown and repair of the multi-grip hands. Participants had access to temporary hands, though it was sometimes not possible to conduct the tests, as the temporary hands also required repairs at times. The authors concluded that the multi-grip myoelectric prosthetic hand can be more useful than a single grip model for performance of specific, individually chosen activities, however the poor durability of the multi-grip hand was problematic for daily use. A durable single-grip myoelectric prosthetic hand may still be needed for heavier physical activities. With structured training, a standard 2-site electrode control system can be used to operate a multi-grip myoelectric prosthetic hand. Further studies are needed into the user's perspective of using multi-grip hands in daily life, to provide valuable information to both clinicians and prosthetic developers.

In a cross-sectional study, (Wanamaker 2019) investigated changes in upper limb function and kinematics in 10 men with partial-hand amputations performing a functional assessment by comparing results with and without a multi-articulating hand prosthesis. Five (5) participants with four-digit loss (thumb intact) and five (5) participants with five-digit loss were included in the study. Three-dimensional kinematics were collected as the participants performed a functional assessment using the Southampton Hand Assessment Procedure (SHAP) with and without a prosthesis. All participants had the i-digits prosthesis from Touch Bionics by Össur. The i-digits is an externally powered partial hand prosthesis with individual articulating digits. Significantly larger joint motions were seen without the prosthesis than with for all participants, which may be an indicator of higher risk for overuse injury. Significant improvement was seen in SHAP scores in the five-digit limb loss participants using the prosthesis compared with not using the device ($p < 0.05$ for 6 of 7 SHAP score categories) though there were no statistical differences in SHAP scores for those with four-digit limb loss. The prosthesis reduced functional deficits and decreased joint range of motion in individuals with partial hand loss. Results showed reduced compensatory motions throughout the upper limb and torso which may reduce the risk of overuse injury. Limitations of the study include the small sample size, the inclusion/exclusion criteria may have resulted in the recruitment of highly functional four-digit limb loss participants and a cohort with good control and strength of the residual limb and may not be representative of the general population. In addition, only one specific prosthesis was investigated, limiting the generalizability to other prostheses for individuals with partial hand loss.

Whelan and Farley (2018) published findings of a case series evaluating an externally powered partial hand prosthesis and the ability to complete functional tasks. Fifteen individuals (12 males) with four- (with thumb remaining) or five-digit partial hand limb loss or absence were fit with the multi-articulating, externally powered i-digits partial hand prostheses. The participants were evaluated using the Southampton Hand Assessment Procedure (SHAP) and Patient-Specific Functional Scale (PSFS). All participants demonstrated a clinically significant change in scores on both the PSFS and

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 14 of 24

the SHAP. The individuals with five-digit absence demonstrated marked improvement in comparison with those with four-digit absence; however, both were far superior to the minimal detectable change score for the SHAP, with 42.33 and 19.16 average improvement scores, respectively. The results indicate that use of the partial hand prosthesis demonstrated functional improvements in objective hand function and individualized goals. The small sample size of 15 participants represents mostly young men with acquired amputations, which may not represent the entire partial-hand population. The individual's remaining range of motion, strength, and function in the thumb and wrist were not fully accounted as potentially influencing results between subjects.

Luke Arm

For individuals who have a missing limb at the wrist or higher who receive sensor and myoelectric controlled upper-limb prosthetic components, the evidence includes a series of publications from a 12-week home study (Resnik 2017, 2017, 2018). Relevant outcomes are functional outcomes and quality of life. The prototypes for the advanced prosthesis were evaluated by the U.S. military and Veterans Administration. Demonstration of improvement in function has been mixed. After several months of home use, activity speed was shown to be similar to the conventional prosthesis, and there were improvements in the performance of some activities, but not all. There were no differences between the prototype and the participants' prostheses for outcomes of dexterity, prosthetic skill, spontaneity, pain, community integration, or quality of life. Study of the current generation of the sensor and myoelectric controlled prosthesis is needed to determine whether newer models of this advanced prosthesis lead to consistent improvements in function and quality of life. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome. The use of such devices has been limited due to their weight, complexity, and learning curve, and are not generally accepted in clinical practice at this time.

Osseointegrated

A retrospective observational cohort study (Muderis 2024) reports the changes in mobility, quality of life (QOL), and the safety profile of the largest described cohort of patients with unilateral Trans-Tibial osseointegration (TTOI) following traumatic amputation. The cohort consisted of patients with data outcomes collected before and after osseointegration intervention. Participants included twenty-one skeletally mature adults who had failed socket prosthesis rehabilitation, with at least two years of post-osseointegration follow-up. All patients demonstrated statistically significant improvement post osseointegration surgery though three (3) patients had four (4) unplanned surgeries: two (2) soft tissue refashioning, and one (1) soft tissue debridement followed eventually by implant removal. No deaths, postoperative systemic complications, more proximal amputations, or periprosthetic fractures occurred. The authors concluded TTOI is likely to confer mobility and quality of life improvements to patients dissatisfied with traditional socket prosthesis rehabilitation following unilateral traumatic transtibial amputation. Adverse events were found to be infrequent and not further disabling. This study is limited by its small sample size. The authors acknowledge the study is not large enough to definitively understand TTOI outcomes or fully establish definitive indications or contraindications for osseointegrated reconstruction. Further prospective study of transtibial osseointegration is needed.

A retrospective analysis (Black 2023) was performed on all patients who underwent single-stage lower limb osseointegration at the single study site institution between 2017 and 2021. Patient demographics, medical history, operative data, and outcomes were collected. Sixty patients met the

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 15 of 24

study criteria: 42 males and 18 females with 35 transfemoral and 25 transtibial amputations. The cohort had an average age of 48 years (range, 25–70 years) and follow-up period of 22 months (range, 6–47 months). Indications for amputation were trauma (50), prior surgical complication (5), cancer (4), and infection (1). Postoperatively, 25 patients developed soft tissue infections, five (5) developed osteomyelitis, six (6) had symptomatic neuromas, and seven (7) required soft tissue revisions. Soft tissue infections were positively correlated with obesity and female sex. Neuroma development was associated with increased age at osseointegration. Neuromas and osteomyelitis were both associated with decreased center experience. Notably, hypertension (15), tobacco use (27), and prior site infection (23) did not correlate with worse outcomes. Forty-seven percent of soft tissue infections occurred in the one (1) month after implantation, and 76% occurred in the first 4 months. The authors concluded the data provides preliminary insights into risk factors for postoperative complications arising from lower limb osseointegration. These factors are both modifiable (body mass index, center experience), and unmodifiable (sex, age). Such results are necessary to inform best practice guidelines and optimize outcomes though further prospective studies are needed.

In a nonrandomized, prospective cohort study (Hagberg 2022), patients with TFAs treated between 1999 and 2007 with the Osseointegrated Prosthesis for the Rehabilitation of Amputees (OPRA) system (n = 51) (28 men/23 women; mean age at amputation: 32 years old; mean age at treatment: 44 years old) in a single university hospital, were followed for 10 years. The objectives were to determine patient-reported outcomes (PROs), and complications after 10 years compared to before treatment and to compare the first five (5) year period with the later five (5) year period with regard to the outcomes. PROs showed statistically significant mean improvements between baseline and the 10-year follow-up. No PROs showed a statistically significant deterioration. Over the 10 years, 12 patients were lost (one (1) lost to follow-up, one (1) dropped out of the study, two (2) died, and eight (8) had implants removed (four (4) before five (5) years and four (4) between five (5) and 10 years). At 10 years, the revision-free survival rates were 83% (CI: 69%-91%), 65% (CI: 49%-77%) and 17% (CI: 7%-29%) for implant revision, deep infection, and mechanical complications, respectively. Mechanical complications, 3.9 per 10 person-years (CI: 2.2-5.1) constituted the most common serious adverse event and were more common during the last five (5) years than during the first five (5) years ($p < 0.001$). No significant difference in the incidence of deep infections was observed between the earlier and the later five (5) year periods: 0.3 per five (5) person-years (CI: 0.1-0.5) vs. 0.3 per person-years (CI: 0.1-0.5) ($p = 0.740$). Correlation analyses between the earlier and later five (5) years revealed a positive association between deep infections and implant removal (0.57, $p < 0.001$) and between mechanical complications and adverse events (0.65, $p < 0.001$). The authors discuss limitations of the study include the number of superficial infections. This data should be interpreted with caution as this adverse event did not need to be diagnosed or treated at the study hospital, and information might be lacking in the records. Other limitations of the study include the lack of a comparable control, small sample size, mixture of patients having both unilateral and bilateral TFAs, and the absence of systematically registered prosthetic device details (i.e., type of prosthetic knee and foot components). Finally, this 10-year follow-up did not include details about other complications commonly reported among individuals living for decades with a lower-limb amputation, such as low back pain, phantom limb pain, falls, and arthrosis in the lower extremity. The authors concluded that improved PROs were demonstrated 10 years after the introduction of a

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 16 of 24

bone anchorage of amputation prostheses. Nevertheless, an increasing rate of mechanical complications is of concern and further clinical study is warranted.

PROFESSIONAL GUIDELINE(S)

In 2024, clinical practice guidelines published by US Department of Veterans Affairs and US Department of Defense regarding rehabilitation of individuals with lower limb amputation recommended offering microprocessor knee units over non-microprocessor knee units for ambulation to reduce risk of falls and maximize patient satisfaction. There is insufficient evidence to recommend for or against any particular socket design, prosthetic foot categories, and suspensions and interfaces. For prosthetic ambulators, they suggest energy storing and return (ESAR) or microprocessor-controlled foot and ankle components over solid ankle cushioned heel (SACH) feet to improve ambulation and patient satisfaction. Newly added with this update, for patients with transfemoral amputation who meet eligibility criteria, they suggest osseointegration as an option to improve prosthesis use based on a weak strength of evidence.

In 2022, clinical practice guidelines published by US Department of Veterans Affairs and US Department of Defense regarding management of upper limb amputation rehabilitation recommend the use of body-powered or externally powered prosthesis to improve independence and reduce disability for patients with major unilateral upper limb amputation (i.e., through or proximal to the wrist). They state there is insufficient evidence to recommend for or against any specific control strategy, socket design, suspension method or component. There is insufficient evidence to recommend for or against the use of any particular recent treatment advances including hardware, software, surgical, technology, or supplemental surgical interventions, such as: targeted muscle reinnervation (TMR), regenerative peripheral nerve interfaces (RPNI), vascularized composite allotransplantation (VCA), agonist-antagonist myoneural interface (AMI), implantable myoelectric sensor system (IMES), or osseointegration (OI).

The National Institute for Health and Care Excellence (NICE) published a health tech guidance for the direct skeletal fixation of limb prostheses using an intraosseous transcutaneous implant (NICE 2024). They state that using an intraosseous transcutaneous implant can only be done currently using special arrangements for clinical governance, consent, and audit or research. They highlight the possible benefits to quality of life and normal daily activities but recognize the evidence for osseointegration is limited in quality and mainly from observational studies. There is currently evidence of serious complications from osseointegration including fractures and infection leading to complications and additional procedures with the possibility to impact mental health as well.

REGULATORY STATUS

Available myoelectric devices include, but are not limited to, ProDigits and i-limb (Touch Bionics), the SensorHand Speed and Michelangelo Hand (Otto Bock), the LTI Boston Digital Arm System (Liberating Technologies), the Utah Arm Systems (Motion Control), and bebionic (Ottobock).

In 2014, the DEKA Arm System (DEKA Integrated Solutions, now DEKA Research & Development), now called the LUKE Arm (Mobius Bionics), was cleared for marketing by FDA through the de novo 513(f)(2) classification process for novel low- to moderate-risk medical devices that are first-of-a-kind.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 17 of 24

In 2015, The OPRA Implant System was approved by the Food and Drug administration (FDA) via humanitarian device exemption. It is indicated for patients who have trans-femoral amputation due to trauma or cancer and who have rehabilitation problems with, or cannot use, a conventional socket prosthesis. The effectiveness of this device for this use has not been demonstrated.

The U.S. Food and Drug Administration (FDA) regulates orthotics as medical devices. All orthotics including related components require FDA approval before marketing and use in the United States to ensure they are safe and effective for human use. Refer to the FDA Medical Device website. Available from: <https://www.fda.gov/medical-devices> [accessed 2026 Apr 16]

The FDA lists the most serious type of medical device recalls as well as early alert communications about corrective actions being taken by companies that the FDA believes are likely to be the most serious type of recalls. Available from: [Medical Device Recalls | FDA](#) [accessed 2026 Apr 16]

CODE(S)

- Codes may not be covered under all circumstances.
- Code list may not be all inclusive (AMA and CMS code updates may occur more frequently than policy updates).
- (E/I)=Experimental/Investigational
- (NMN)=Not medically necessary/appropriate

CPT Codes

Code	Description
Not Applicable	

Copyright © 2026 American Medical Association, Chicago, IL

HCPCS Codes

Code	Description
A4361-A4437	Ostomy supplies (code range)
A5051-A5093	Additional ostomy supplies (code range)
A9282	Wig, any type, each
L3250	Orthopedic footwear, custom molded shoe, removable inner mold, prosthetic shoe, each
L5000-L5855	Lower limb prosthetic (code range)
L5615	Addition, endoskeletal knee-shin system, 4 bar linkage or multiaxial, fluid swing and stance phase control
L5827	Endoskeletal knee-shin system, single axis, electromechanical swing and stance phase control, with or without shock absorption and stance extension damping

Medical Policy: External Prosthetic Devices**Policy Number: 1.01.18****Page: 18 of 24**

Code	Description
L5856	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control knee feature, swing, and stance phase, includes electronic sensor(s), any type
L5857	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control feature, swing phase only, includes electronic sensor(s), any type
L5858	Addition to lower extremity prosthesis, endoskeletal knee-shin system, microprocessor control feature, stance phase only, includes electronic sensor(s), any type
L5859	Addition to lower extremity prosthesis, endoskeletal knee-shin system, powered and programmable flexion/extension assist control, includes any type motor(s)
L5910-L5972	Lower limb prosthetic (code range)
L5969 (NMN)	Addition, endoskeletal ankle-foot or ankle system, power assist, includes any type motor(s)
L5973	Endoskeletal ankle foot system, microprocessor-controlled feature, dorsiflexion, and/or plantar flexion control, includes power source
L5974-L5999	Lower limb prosthetic (code range)
L5991 (E/I)	Addition to lower extremity prostheses, osseointegrated external prosthetic connector
L6000-L6020	Upper limb prosthetic device (code range)
L6026 (E/I)	Transcarpal/metacarpal or partial hand disarticulation prosthesis, external power, self-suspended, inner socket with removable forearm section, electrodes, and cables, two batteries, charger, myoelectric control of terminal device, excludes terminal device(s)
L6028	Partial hand including fingers, flexible or non-flexible interface, endoskeletal system, molded to patient model, for use without external power, not including inserts described by L6692
L6029	Upper extremity addition, test socket/interface, partial hand including fingers
L6030	Upper extremity addition, external frame, partial hand including fingers
L6031	Replacement socket/interface, partial hand including fingers, molded to patient model, for use with or without external power

Medical Policy: External Prosthetic Devices**Policy Number: 1.01.18****Page: 19 of 24**

Code	Description
L6032	Addition to upper extremity prosthesis, partial hand including fingers, ultralight material (titanium, carbon fiber or equal)
L6033	Addition to upper extremity prosthesis, partial hand including fingers, acrylic material
L6034	Partial hand, finger, and thumb prosthesis without prosthetic digit(s)/thumb, amputation at transmetacarpal level, including flexible or non-flexible interface, molded to patient model, for use without external power and/or passive prosthetic digit/thumb, not including inserts described by L6692
L6035	Single prosthetic digit, mechanical, can include metacarpophalangeal (MCP), proximal interphalangeal (PIP), and/or distal interphalangeal (DIP) joint(s), with or without locking mechanism, can include flexion or extension assist, any material, attachment, initial issue or replacement
L6036	Prosthetic thumb, mechanical, can include metacarpophalangeal (MCP), interphalangeal (IP) joint(s), with or without locking mechanism, can include flexion or extension assist, any material, attachment, initial issue or replacement
L6037	Immediate post-surgical or early fitting, application of initial rigid dressing, including fitting alignment and suspension of components, and one cast change, partial hand including fingers
L6038	Addition to single prosthetic digit or thumb, mechanical, attachment, multiaxial and/or internal/external rotation/abduction/adduction mechanism, with or without locking feature, any material
L6039	Passive prosthetic digit or thumb prosthesis not including hand restoration partial hand, full or partial, custom made, any material, initial or replacement, per single passive prosthetic digit or thumb
L6050-L6698	Upper limb prosthetic device (code range)
L6700	Upper extremity addition, external powered feature, myoelectronic control module, additional emg inputs, pattern-recognition decoding intent movement
L6703-L6810	Terminal devices (hooks) (code range)
L6880 (E/I)	Electric hand, switch or myoelectric controlled, independently articulating digits, any grasp pattern or combination of grasp patterns, includes motor(s)
L6881	Automatic grasp feature, addition to upper limb electric prosthetic terminal device
L6882	Microprocessor control feature, addition to upper limb prosthetic terminal device

Medical Policy: External Prosthetic Devices**Policy Number: 1.01.18****Page: 20 of 24**

Code	Description
L6883-L6885	Replacement socket (code range)
L6890-L6915	Hand – gloves – hand restoration (code range)
L6920	Wrist disarticulation, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal switch, cables, two batteries and one charger, switch control of terminal device
L6925 (E/I)	Wrist disarticulation, external power, self-suspended inner socket, removable forearm shell, Otto Bock or equal electrodes, cables, two batteries and one charger, myoelectronic control of terminal device
L6930-L6975	External power (base devices) (code range)
L7259	Electronic wrist rotator, any type
L7360-L7368	Battery components (code range)
L7406	Addition to upper extremity, user adjustable, mechanical, residual limb volume management system
L7499	Upper extremity prosthesis, not otherwise specified
L7510	Repair of prosthetic device, repair or replace minor parts
L7520	Repair prosthetic device, labor component, per 15 minutes
L7700	Gasket or seal, for use with prosthetic socket insert, any type, each
L8000–L8039	General prosthesis; breast (code range)
L8040-L8049	General prosthesis; face and ear (code range)
L8400-L8499	Prosthetic socks (shrinker, sheath, stump sock) (code range)
L8500–L8515	Larynx and trachea prosthetics and accessories (code range)
L9900	Orthotic and prosthetic supply, accessory, and/or service component of another HCPCS L code
V2623-V2629	Prosthesis, ocular (code range)

ICD10 Codes

Code	Description
Multiple Codes	

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 21 of 24

REFERENCES

Agrawal V, et al. Comparison of four different categories of prosthetic feet during ramp ambulation in unilateral transtibial amputees. *Prosthet Orthot Int*. 2015 Oct;39(5):380-9.

Alzeer AM, et al. Impacts of microprocessor-controlled versus non-microprocessor-controlled prosthetic knee joints among transfemoral amputees on functional outcomes: a comparative study. *Cureus Apr*. 2022; 14(4): e24331.

Black GG, et al. Osseointegration for lower limb amputation: understanding the risk factors and time courses of soft tissue complications. *Ann Plast Surg*. 2023 Jun 1;90(6S Suppl 5):S452-S456.

Colas-Ribas C, et al. Effects of a microprocessor-controlled ankle-foot unit on energy expenditure, quality of life, and postural stability in persons with transtibial amputation: An unblinded, randomized, controlled, cross-over study. *Prosthet Orthot Int*. 2022 Dec;46(6):541-548.

Darter BJ, et al. Energetic consequences of using a prosthesis with adaptive ankle motion during slope walking in persons with a transtibial amputation. *Prosthet Orthot Int*. 2014; 38(1):5-11.

Department of Veterans Affairs; Department of Defense [Internet]. VA/DoD clinical practice guideline for rehabilitation of individuals with lower limb amputation. Version 3.0. Washington (DC): U.S. Government Printing Office; 2024 Dec [accessed 2026 Apr 16]. Available from: https://www.healthquality.va.gov/HEALTHQUALITY/guidelines/Rehab/amp/LLA-CPG_2024-Guideline_final_20250110.pdf

Department of Veterans Affairs; Department of Defense [Internet]. VA/DoD clinical practice guideline for the management of upper limb amputation rehabilitation. Version 2.0. Washington (DC): U.S. Government Printing Office; 2022 Mar [accessed 2026 Apr 16]. Available from: https://www.healthquality.va.gov/HEALTHQUALITY/guidelines/Rehab/ULA/VADODULACPG_Final_508.pdf

Eberly VJ, et al. Impact of a stance phase microprocessor-controlled knee prosthesis on level walking in lower functioning individuals with a transfemoral amputation. *Prosthet Orthot Int*. 2014 Dec;38(6):447-55.

Ferris AE, et al. Evaluation of a powered ankle-foot prosthetic system during walking. *Arch Phys Med Rehabil*. 2012;93(11):1911-8.

Gailey RS, et al. Application of self-report and performance-based outcome measures to determine functional differences between four categories of prosthetic feet. *J Rehabil Res Dev*. 2012;49(4):597-612.

Hagberg K, et al. Osseointegrated prostheses for the rehabilitation of patients with transfemoral amputations: A prospective ten-year cohort study of patient-reported outcomes and complications. *J Orthop Translat*. 2022 Oct 20;38:56-64.

Hahn A, et al. The effect of microprocessor controlled exo-prosthetic knees on limited community ambulators: systematic review and meta-analysis. *Disabil Rehabil*. 2022 Dec; 44(24): 7349-7367.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 22 of 24

Herr HM, et al. Bionic ankle-foot prosthesis normalizes walking gait for persons with leg amputation. *Proc Biol Sci.* 2012;279(1728):457-64.

Highsmith MJ, et al. Ramp descent performance with the C-Leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int.* 2013 Oct;37(5):362-8.

Johansen D, et al. Hybrid tongue-myoelectric control improves functional use of a robotic hand prosthesis. *IEEE Trans Biomed Eng.* 2021 Jun;68(6):2011-2020.

Kaufman KR, et al. Gait asymmetry of transfemoral amputees using mechanical and microprocessor-controlled prosthetic knees. *Clin Biomech (Bristol).* 2012 Jun;27(5):460-5.

Kerver N, et al. The multi-grip and standard myoelectric hand prosthesis compared: does the multi-grip hand live up to its promise? *J Neuroeng Rehabil.* 2023 Feb 15;20(1):22.

Li S, et al. Physiological parameters analysis of transfemoral amputees with different prosthetic knees. *Acta Bioeng Biomech.* 2019;21(3):135-142.

Muderis MA, et al. Transtibial osseointegration following unilateral traumatic amputation: An observational study of patients with at least two years follow-up. *Injury.* 2024 Jun;55(6):111568.

National Institute for Health and Care Excellence (NICE) [Internet]. Healthtech guidance: Direct skeletal fixation of limb prostheses using an intraosseous transcutaneous implant (HTG732). 2024 Dec 17 [accessed 2026 Apr 16]. Available from: <https://www.nice.org.uk/guidance/htg732>

Ontario Health (Quality). Osseointegrated prosthetic implants for people with lower-limb amputation: a health technology assessment. *Ont Health Technol Assess Ser.* 2019 Dec 12;19(7):1-126.

Resnik L, et al. Does the DEKA Arm substitute for or supplement conventional prostheses. *Prosthet Orthot Int.* 2018 Oct; 42(5): 534-543.

Resnik LJ, et al. How do the outcomes of the DEKA Arm compare to conventional prostheses?. *PLoS One.* 2018 Jan 17; 13(1): e0191326.

Resnik LJ, et al. Perceptions of satisfaction, usability, and desirability of the DEKA Arm before and after a trial of home use. *PLoS One.* 2017 Jun 2;12(6):e0178640.

Schmalz T, et al. Energy expenditure and biomechanical characteristics of lower limb amputee gait: the influence of prosthetic alignment and different prosthetic components. *Gait Posture.* 2002;16:255-63.

Theeven P, et al. Functional added value of microprocessor-controlled knee joints in daily life performance of Medicare Functional Classification Level-2 amputees. *J Rehabil Med.* 2011; 43(10):906-15.

Theeven PJ, et al. Influence of advanced prosthetic knee joints on perceived performance and everyday life activity level of low-functional persons with a transfemoral amputation or knee disarticulation. *J Rehabil Med.* 2012; 44(5):454-61.

Thibaut A, et al. Impact of microprocessor prosthetic knee on mobility and quality of life in patients with lower limb amputation: a systematic review of the literature. *Eur J Phys Rehabil Med.* Jun 2022; 58(3): 452-461.

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 23 of 24

Thomas-Pohl M, et al. Microprocessor prosthetic ankles: comparative biomechanical evaluation of people with transtibial traumatic amputation during standing on level ground and slope. Disabil Rehabil Assist Technol. Jan 2021; 16(1): 17-26.

Wanamaker AB, et al. Biomechanical analysis of users of multi-articulating externally powered prostheses with and without their device. Prosthet Orthot Int. 2019 Dec; 43(6):618-628.

Whelan, L, et al. Functional outcomes with externally powered partial hand prostheses. Jour Pros and Orth. 2018 Apr; 30(2):p 69-73.

Widehammar C, et al. Effect of multi-grip myoelectric prosthetic hands on daily activities, pain-related disability and prosthesis use compared with single-grip myoelectric prostheses: A single-case study. J Rehabil Med. 2022 Jan 3;54:jrm00245.

SEARCH TERMS

Not Applicable

CENTERS FOR MEDICARE AND MEDICAID SERVICES (CMS)

[Prosthetic Shoe \(NCD 280.10\)](#) [accessed 2026 Apr 16]

[External Breast Prostheses \(LCD L33317\)](#) [accessed 2026 Apr 16]

[Eye Prostheses \(LCD L33737\)](#) [accessed 2026 Apr 16]

[Facial Prostheses \(LCD L33738\)](#) [accessed 2026 Apr 16]

[Lower Limb Prostheses \(LCD L33787\)](#) [accessed 2026 Apr 16]

Upper Extremity Prostheses is not addressed in National or Regional Medicare coverage determinations or policies.

PRODUCT DISCLAIMER

- Services are contract dependent; if a product does not cover a service, medical policy criteria do not apply.
- If a commercial product (including an Essential Plan or Child Health Plus product) covers a specific service, medical policy criteria apply to the benefit.
- If a Medicaid product covers a specific service, and there are no New York State Medicaid guidelines (eMedNY) criteria, medical policy criteria apply to the benefit.
- If a Medicare product (including Medicare HMO-Dual Special Needs Program (DSNP) product) covers a specific service, and there is no national or local Medicare coverage decision for the service, medical policy criteria apply to the benefit.
- If a Medicare HMO-Dual Special Needs Program (DSNP) product DOES NOT cover a specific service, please refer to the Medicaid Product coverage line.

POLICY HISTORY/REVISION

Medical Policy: External Prosthetic Devices

Policy Number: 1.01.18

Page: 24 of 24

Committee Approval Dates	
07/25/02, 10/23/03, 05/27/04, 04/28/05, 04/27/06, 04/26/07, 02/26/09, 02/25/10, 06/24/11, 08/23/12, 06/27/13, 06/26/14, 02/26/15, 02/25/16, 02/16/17, 02/22/18, 02/28/19, 02/27/20, 02/25/21, 02/17/22, 02/16/23, 02/22/24, 04/17/25, 04/16/26	
Date	Summary of Changes
10/28/26	<ul style="list-style-type: none">Off-cycle policy edit, E/I notation removed from L6881 and L6882; policy intent unchanged.
04/16/26	<ul style="list-style-type: none">Annual review; policy statement added for microprocessor-controlled ankle-foot prosthetics as medically necessary and policy statements added for repair and replacement of prosthetic devices.
09/30/25	<ul style="list-style-type: none">Off-cycle policy review, code edit, added HCPCS codes L5657, L6034-L6036, L6038 and L6039. Policy intent unchanged.
04/17/25	<ul style="list-style-type: none">Annual review, policy statement added for prosthetics not for ADLs as not medically necessary, osseointegrated, myoelectric partial hand, and enhanced dexterity upper limb prosthetic devices as investigational, revision to Policy Guidelines, code edit, added L5991.
01/01/25	<ul style="list-style-type: none">Summary of changes tracking implemented.
07/25/02	<ul style="list-style-type: none">Original effective date